

# Crowcon Technical Note

**Document Reference:** GEN089 Sensor Cross Sensitivity (Portable Instruments)  
**Document applies to:** Crowcon Portable Instruments  
**Release Date:** 06/03/2019  
**Subject:** Sensor cross sensitivity guidance and warnings

## Sensor Cross Sensitivity Guidance, Advice and Warnings

### Introduction:

Toxic sensors are designed and built to demonstrate an optimum response to the target gas. However, interfering gases in the environment can affect the reading given. The information contained in this note has been collated to offer guidance and explain these cross sensitivities.

Crowcon regularly review the sensors available in the market, and focus on using the best possible solutions, many with active filters that assist in both reducing the possibility of inhibiting, poisoning and limiting cross sensitive response.

Flammable sensors are commonly available in two options, pellistor (catalytic) or InfraRed (IR). Whilst the pellistor option is recognised as lower cost and suited to most applications, it is susceptible to poisoning (silicone and hydrogen sulphide), is not failsafe and can be damaged following exposure to greater than 100% LEL levels of gas. Most Crowcon portables offer the 'pellistor saver' mode, which actively turns off the sensor whilst at the same time notifying the user that the sensor is not currently enabled (the monitor should at this point be in alarm as the saver mode is commonly set at 90% LEL).

Infrared sensors are currently available for flammable or carbon dioxide detection. IR options offer failsafe detection, lower battery consumption (meaning that portable monitors will run for longer from one charge) and immunity to potential poisons.

The most popular technology for measuring oxygen concentration in air is the galvanic oxygen sensor. As the lead is oxidised, the lifetime of the sensor is limited to either 2 or 3 years. Oxygen sensors are not commonly inhibited or cross sensitive to other gases.

### What is cross sensitivity?

There are a number of forms of cross sensitivity, and more than one can be relevant at one time:

1. Interference readings causing false alarms or no alarm
  - i. Positive response to a non target gas, causes a positive reading
  - ii. Negative response to a non target gas, causes a negative reading
2. Inhibited response for a timeframe after exposure without causing long term damage
3. Poisoning which renders the sensor inoperable

Cross sensitivities can differ depending on the ambient temperature and level of gas the sensor is being exposed to. The information below is provided as a guide, and a thorough risk assessment to understand the potential gases in the environment is always recommended.

# Crowcon Technical Note

## Proving Response:

The best way to ensure the correct response of any sensor is to perform a regular gas test (bump test). Whilst there are increasingly strong recommendations for testing flammable sensors (EN60079-29 parts 1 and 2) the testing of toxic gases is not yet formalised. Most Crowcon offer this capability; please review the individual manuals for details.

## Toxic gas sensors and their cross sensitivity response

The following table provides a guide to sensor cross sensitivities, and should be viewed with the 'Cross Sensitivity Chart' for easy application. Only sensors with cross sensitivity issues are listed.

Sensor	Common 1 <sup>st</sup> Alarm	Cross Sensitivity	Description / Recommended action	Status
Ammonia (NH <sub>3</sub> )	25 ppm	Hydrogen sulphide H <sub>2</sub> S	Low positive response, this is a safe situation as the NH <sub>3</sub> sensor will show a positive reading in the presence of H <sub>2</sub> S. In extreme cases this could cause a false alarm.	
Ammonia (NH <sub>3</sub> )	25 ppm	Sulphur Dioxide SO <sub>2</sub>	Low to medium negative response	
Arsine AsH <sub>3</sub>	0.05 ppm	Hydrogen Chloride HCl	Low positive response, this is a safe situation as the AsH <sub>3</sub> sensor will show a positive reading in the presence of HCL.	
Arsine AsH <sub>3</sub>	0.05 ppm	Hydrogen Cyanide HCN	Low positive response, this is a safe situation as the AsH <sub>3</sub> sensor will show a positive reading in the presence of HCN.	
Arsine AsH <sub>3</sub>	0.05 ppm	Hydrogen Sulphide H <sub>2</sub> S	Medium to high positive response, this is a safe situation as the AsH <sub>3</sub> sensor will show a positive reading in the presence of H <sub>2</sub> S.	
Arsine AsH <sub>3</sub>	0.05 ppm	Nitrogen dioxide NO <sub>2</sub>	Negative reading.	
Arsine AsH <sub>3</sub>	0.05 ppm	Phosphine PH <sub>3</sub>	High positive response, this is a safe situation as the AsH <sub>3</sub> sensor will show a positive reading in the presence of PH <sub>3</sub> .	
Arsine AsH <sub>3</sub>	0.05 ppm	Sulphur Dioxide SO <sub>2</sub>	Medium positive response, this is a safe situation as the AsH <sub>3</sub> sensor will show a positive reading in the presence of SO <sub>2</sub> .	
Carbon Monoxide CO	30 ppm	Acetylene C <sub>2</sub> H <sub>2</sub>	Unsafe response, CO sensor saturated at <2.5%LEL C <sub>2</sub> H <sub>2</sub> .	
Carbon Monoxide CO	30 ppm	Ethylene C <sub>2</sub> H <sub>4</sub>	Unsafe response, CO sensor saturated at <2.5%LEL C <sub>2</sub> H <sub>4</sub> .	
Carbon Monoxide CO	30 ppm	Ethanol C <sub>2</sub> H <sub>6</sub> O	Medium positive response, false alarms possible. Tests have shown that CO sensor will recover from short term exposure to low %LEL levels of ethanol.	
Carbon Monoxide CO	30 ppm	Hydrogen H <sub>2</sub>	High positive response. The H <sub>2</sub> filtered sensor is therefore recommended for this combination.	
Carbon Monoxide CO	30 ppm	Nitric oxide NO	Low positive response, this is a safe situation as the CO sensor will show a positive reading in the presence of NO. In extreme cases this could cause a false alarm. The H <sub>2</sub> filtered sensor does not exhibit this response and is therefore recommended for this combination where at all possible.	
Chlorine Cl <sub>2</sub>	0.3 ppm	Chlorine dioxide ClO <sub>2</sub>	Medium positive response, this is a safe situation as the Cl <sub>2</sub> sensor will show a positive reading in the presence of ClO <sub>2</sub> . In extreme cases this could cause a false alarm.	

# Crowcon Technical Note

Sensor	Common 1 <sup>st</sup> Alarm	Cross Sensitivity	Description / Recommended action	Status
Chlorine Cl <sub>2</sub>	0.3 ppm	Fluorine F <sub>2</sub>	Medium positive response, this is a safe situation as the Cl <sub>2</sub> sensor will show a positive reading in the presence of F <sub>2</sub> .	
Chlorine Cl <sub>2</sub>	0.3 ppm	Hydrogen sulphide H <sub>2</sub> S	Reduce H <sub>2</sub> S alarm level to 2.5ppm and check response of Cl <sub>2</sub> sensor after alarm event. The Cl <sub>2</sub> sensor will recover quickly at these low levels, reducing the chance of permanent damage.	
Chlorine Cl <sub>2</sub>	0.3 ppm	Nitrogen dioxide NO <sub>2</sub>	Medium positive response, this is a safe situation as the Cl <sub>2</sub> sensor will show a positive reading in the presence of NO <sub>2</sub> .	
Chlorine Cl <sub>2</sub>	0.3 ppm	Ozone O <sub>3</sub>	Medium positive response, this is a safe situation as the Cl <sub>2</sub> sensor will show a positive reading in the presence of O <sub>3</sub> .	
Chlorine dioxide ClO <sub>2</sub>	0.1 ppm	Hydrogen sulphide H <sub>2</sub> S	Negative response at even very low levels, this also inhibits the sensor for a significant amount of time.	
Chlorine dioxide ClO <sub>2</sub>	0.1 ppm	Chlorine Cl <sub>2</sub>	High positive response, this is a safe situation as the ClO <sub>2</sub> sensor will show a positive reading in the presence of Cl <sub>2</sub> .	
Chlorine dioxide ClO <sub>2</sub>	0.1 ppm	Ozone O <sub>3</sub>	High positive response, this is a safe situation as the ClO <sub>2</sub> sensor will show a positive reading in the presence of O <sub>3</sub> .	
Ethylene Oxide	5 ppm	Carbon Monoxide CO	Medium positive response, this is a safe situation as the ETO sensor will show a positive reading in the presence of CO.	
Fluorine F <sub>2</sub>	1 ppm	Arsine AsH <sub>3</sub>	Low to medium negative response	
Fluorine F <sub>2</sub>	1 ppm	Chlorine Cl <sub>2</sub>	High positive response, this is a safe situation as the F <sub>2</sub> sensor will show a positive reading in the presence of Cl <sub>2</sub> .	
Fluorine F <sub>2</sub>	1 ppm	Hydrogen Chloride HCl	Medium to high negative response	
Fluorine F <sub>2</sub>	1 ppm	Hydrogen Sulphide H <sub>2</sub> S	Medium to high negative response	
Fluorine F <sub>2</sub>	1 ppm	Nitrogen Dioxide NO <sub>2</sub>	High positive response, this is a safe situation as the F <sub>2</sub> sensor will show a positive reading in the presence of NO <sub>2</sub> .	
Fluorine F <sub>2</sub>	1 ppm	Ozone O <sub>3</sub>	High positive response, this is a safe situation as the F <sub>2</sub> sensor will show a positive reading in the presence of O <sub>3</sub> .	
Fluorine F <sub>2</sub>	1 ppm	Phosphine	Medium negative response	
Fluorine F <sub>2</sub>	1 ppm	Sulphur Dioxide SO <sub>2</sub>	Low to medium negative response	
Hydrogen H <sub>2</sub> 2000ppm range	999ppm	Carbon Monoxide CO	Medium positive response, this is a safe situation as the H <sub>2</sub> sensor will show a positive reading in the presence of CO.	
Hydrogen H <sub>2</sub> 2000ppm range	999ppm	Nitrogen Monoxide NO	Medium positive response, this is a safe situation as the H <sub>2</sub> sensor will show a positive reading in the presence of NO.	
Hydrogen H <sub>2</sub> 2000ppm range	999ppm	Ethylene C <sub>2</sub> H <sub>4</sub>	High positive response, sensor saturated at <1%LEL C <sub>2</sub> H <sub>4</sub> . Not recommended for use in applications with flammable levels of C <sub>2</sub> H <sub>4</sub> are expected.	

# Crowcon Technical Note

Sensor	Common 1 <sup>st</sup> Alarm	Cross Sensitivity	Description / Recommended action	Status
Hydrogen cyanide HCN	0.9 ppm	Nitrogen Monoxide NO	Low negative response	
Hydrogen cyanide HCN	0.9 ppm	Nitrogen Dioxide NO <sub>2</sub>	High negative response	
Hydrogen cyanide HCN	0.9 ppm	Hydrogen Sulphide	After filter saturation there will be a high positive response. This is a safe situation as the HCN sensor will show a positive reading in the presence of H <sub>2</sub> S.	
Hydrogen chloride HCl	1 ppm	Arsine AsH <sub>3</sub>	High positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of AsH <sub>3</sub> .	
Hydrogen chloride HCl	1 ppm	Chlorine Cl <sub>2</sub>	Low level response which can be positive or negative.	
Hydrogen chloride HCl	1 ppm	Hydrogen Cyanide HCN	Medium to high positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of HCN	
Hydrogen chloride HCl	1 ppm	Hydrogen Sulphide H <sub>2</sub> S	High positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of H <sub>2</sub> S.	
Hydrogen chloride HCl	1 ppm	Nitrogen Monoxide NO	Medium to High positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of NO.	
Hydrogen chloride HCl	1 ppm	Nitrogen Dioxide NO <sub>2</sub>	Low level response which can be positive or negative.	
Hydrogen chloride HCl	1 ppm	Phosphine PH <sub>3</sub>	High positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of PH <sub>3</sub> .	
Hydrogen chloride HCl	1 ppm	Sulphur Dioxide	High positive response, this is a safe situation as the HCl sensor will show a positive reading in the presence of SO <sub>2</sub> .	
Hydrogen sulphide H <sub>2</sub> S	5 ppm	Chlorine	Low negative response, 1ppm of Cl <sub>2</sub> (please note the common alarm level is 0.3ppm) will be indicated by a -0.25ppm reading on the H <sub>2</sub> S sensor. Reduce H <sub>2</sub> S alarm level to 2.5ppm.	
Hydrogen sulphide H <sub>2</sub> S	5 ppm	Sulphur dioxide SO <sub>2</sub>	Low positive response, this is a safe situation as the H <sub>2</sub> S sensor will show a positive reading in the presence of SO <sub>2</sub> . In extreme cases this could cause a false alarm.	
Hydrogen sulphide H <sub>2</sub> S	5 ppm	Nitrogen Monoxide NO	Low positive response, this is a safe situation as the H <sub>2</sub> S sensor will show a positive reading in the presence of NO. In extreme cases this could cause a false alarm.	
Hydrogen sulphide H <sub>2</sub> S	5 ppm	Nitrogen dioxide NO <sub>2</sub>	Low negative response. Reduce H <sub>2</sub> S alarm level to 2.5ppm.	
Nitrogen Monoxide NO	2 ppm	Hydrogen sulphide H <sub>2</sub> S	Medium positive response, this is a safe situation as the NO sensor will show a positive reading in the presence of H <sub>2</sub> S. In extreme cases this could cause a false alarm.	
Nitrogen Monoxide NO	2 ppm	Chlorine Cl <sub>2</sub>	Low positive response, this is a safe situation as the NO sensor will show a positive reading in the presence of Cl <sub>2</sub> . In extreme cases this could cause a false alarm.	
Nitrogen Monoxide NO	2 ppm	Sulphur dioxide SO <sub>2</sub>	Low positive response, this is a safe situation as the NO sensor will show a positive reading in the presence of SO <sub>2</sub> . In extreme cases this could cause a false alarm.	

# Crowcon Technical Note


Sensor	Common 1 <sup>st</sup> Alarm	Cross Sensitivity	Description / Recommended action	Status
Nitrogen Monoxide NO	2 ppm	Nitrogen dioxide NO <sub>2</sub>	Low positive response, this is a safe situation as the NO sensor will show a positive reading in the presence of NO <sub>2</sub> . In extreme cases this could cause a false alarm.	Blue
Nitrogen dioxide NO <sub>2</sub>	0.5 ppm	Chlorine Cl <sub>2</sub>	High positive response, this is a safe situation as the NO <sub>2</sub> sensor will show a positive reading in the presence of Cl <sub>2</sub> . In extreme cases this could cause a false alarm.	Blue
Nitrogen dioxide NO <sub>2</sub>	0.5 ppm	Hydrogen sulphide H <sub>2</sub> S	Low negative response. Reduce H <sub>2</sub> S alarm level to 2.5ppm.	Green
Nitrogen dioxide NO <sub>2</sub>	0.5 ppm	Sulphur dioxide SO <sub>2</sub>	Low negative response, this situation should be highlighted to the user as SO <sub>2</sub> on the NO <sub>2</sub> sensor could lead to a low reading. However with EH40 alarm levels in place this is a safe situation as alarms will have sounded prior to incorrect readings.	Blue
Ozone O <sub>3</sub>	0.1 ppm	Chlorine Cl <sub>2</sub>	High positive response, this is a safe situation as the O <sub>3</sub> sensor will show a positive reading in the presence of Cl <sub>2</sub> . In extreme cases this could cause a false alarm.	Blue
Ozone O <sub>3</sub>	0.1 ppm	Chlorine dioxide ClO <sub>2</sub>	High positive response, this is a safe situation as the O <sub>3</sub> sensor will show a positive reading in the presence of ClO <sub>2</sub> . In extreme cases this could cause a false alarm.	Blue
Ozone O <sub>3</sub>	0.1 ppm	Hydrogen sulphide H <sub>2</sub> S	Negative response at even very low levels, this also inhibits the sensor for a significant amount of time.	Red
Ozone O <sub>3</sub>	0.1 ppm	Nitrogen dioxide NO <sub>2</sub>	High positive response, this is a safe situation as the O <sub>3</sub> sensor will show a positive reading in the presence of NO <sub>2</sub> . In extreme cases this could cause a false alarm.	Blue
Phosphine PH <sub>3</sub>	0.1 ppm	Arsine AsH <sub>3</sub>	High positive response, this is a safe situation as the PH <sub>3</sub> sensor will show a positive reading in the presence of AsH <sub>3</sub> .	Blue
Phosphine PH <sub>3</sub>	0.1 ppm	Chlorine Cl <sub>2</sub>	Low to medium negative response.	Red
Phosphine PH <sub>3</sub>	0.1 ppm	Nitrogen Dioxide NO <sub>2</sub>	Low to medium negative response.	Red
Phosphine PH <sub>3</sub>	0.1 ppm	Sulphur Dioxide SO <sub>2</sub>	Low to medium positive response on some sensors, this is a safe situation as the PH <sub>3</sub> sensor would show a positive reading in the presence of SO <sub>2</sub> .	Blue
Phosgene COCl <sub>2</sub>	0.02ppm	Chlorine Cl <sub>2</sub>	Medium to high positive response. This is a safe situation as the COCl <sub>2</sub> sensor will show a positive reading in the presence of Cl <sub>2</sub> .	Blue
Phosgene COCl <sub>2</sub>	0.02ppm	Hydrogen Chloride HCl	High positive response. This is a safe situation as the COCl <sub>2</sub> sensor will show a positive reading in the presence of HCl.	Blue
Phosgene COCl <sub>2</sub>	0.02ppm	Nitrogen Dioxide NO <sub>2</sub>	Low to medium negative response.	Red
Sulphur dioxide SO <sub>2</sub>	2 ppm	Chlorine Cl <sub>2</sub>	Negative reading,	Red
Sulphur dioxide SO <sub>2</sub>	2 ppm	Nitrogen Dioxide NO <sub>2</sub>	Negative reading,	Red
Sulphur dioxide	2 ppm	Ethylene	Low positive response, this is a safe situation as the SO <sub>2</sub> sensor will show a positive reading in the presence of C <sub>2</sub> H <sub>4</sub> . In extreme cases this could cause a false alarm.	Blue

# Crowcon Technical Note

Status Key:

	Disruptive
	Under caution
	Lower Alarm Level

Note: Where CO or H2S are listed the detail is relevant to both single and dual sensors.

		Cross Sensitivity Chart, reference - GEN089																					
Issue 1		Sensor in Crowcon Portable																					
		Oxygen	Ammonia	Arsine	Carbon monoxide	Carbon monoxide (Hz compensated)	Carbon dioxide	Chlorine	Chlorine dioxide	Ethylene Oxide	Fluorine	Hydrogen (2000ppm)	Hydrogen Cyanide	Hydrogen Chloride	Hydrogen sulphide	Nitrogen Monoxide	Nitrogen dioxide	Ozone	Phosphine	Phosgene	Sulphur dioxide		
Pumped		Yes	Yes	No	Yes	Yes	Yes	No	No	No	No	Yes	Yes	No	Yes	No	No	No	No	No	No	Yes	
		O <sub>2</sub>	NH <sub>3</sub>	AsH <sub>3</sub>	CO	CO	CO <sub>2</sub>	Cl <sub>2</sub>	ClO <sub>2</sub>	EIO	F <sub>2</sub>	H <sub>2</sub>	HCN	HCl	H <sub>2</sub> S	NO	NO <sub>2</sub>	O <sub>3</sub>	PH <sub>3</sub>	COCl <sub>2</sub>	SO <sub>2</sub>		
Gas Applied	Oxygen O <sub>2</sub>																						
	Ammonia NH <sub>3</sub>																						
	Arsine AsH <sub>3</sub>																						
	Carbon Monoxide CO																						
	Carbon Dioxide CO <sub>2</sub>																						
	Chlorine Cl <sub>2</sub>		+ Med																				
	Chlorine Dioxide ClO <sub>2</sub>																						
	Ethylene Oxide EIO																						
	Fluorine F <sub>2</sub>																						
	Hydrogen H <sub>2</sub>																						
	Hydrogen Cyanide HCN																						
	Hydrogen Chloride HCl																						
	Hydrogen Sulphide H <sub>2</sub> S																						
	Nitrogen Monoxide NO																						
	Nitrogen Dioxide NO <sub>2</sub>																						
	Ozone O <sub>3</sub>																						
	Phosphine PH <sub>3</sub>																						
Phosgene COCl <sub>2</sub>																							
Sulphur Dioxide SO <sub>2</sub>																							
Ethanol																							
Acetylene C <sub>2</sub> H <sub>2</sub>																							
Ethylene C <sub>2</sub> H <sub>4</sub>																							
Disruptive												Should not be configured in the same detector. If the gas on the left is present in the application caution should be exercised.											
Under caution												Cross sensitivity indication, should the gas on the left be applied to the sensor. (Low=<15%, Med =15 to 60%, High equals >60%)											
Lower Alarm Level												The alarm level of the disruptive gas (shown in the cell) should be lowered to increase safety											
OK												No issues											
Disclaimer: Please note, sensor cross sensitivity is dependant on sensors used and gases in the working environment. Sensitivity may be dependant on exposure level and time. The best way to ensure sensor performance is to regularly bump test and calibrate according to the manufacturers recommendations of use.																							

For more information please contact +44 (0)1235 557711, [customersupport@crowcon.com](mailto:customersupport@crowcon.com)