

# Crowcon Technical Note

**Document Reference: GEN091**

**Document applies to: T3 and Gasman**

**Release Date: 08/07/2019**

**Subject: Servicing and replacement of CO sensors in T3 and Gasman**



We are changing our CO sensor to a new version in our T3 and Gasman portable gas detector models.

Please use the below new part code to order replacement CO sensors for T3 in ATEX regions only and for Gasman in ATEX and INMETRO regions only

**New: S011811/S**

- This applies to existing T3 & Gasman units already in the field as well as new orders
- Calibration procedure and gas requirements and procedures remain the same as before
- Pricing remains the same as the existing version

**\*\*Please note for INMETRO and CSA certified T3 CO sensors and CSA certified Gasman CO sensors you must continue to use the existing part code\*\***

**Existing: S011422/M**

---

For more information please contact +44 (0)1235 557711, [technicalsupport@crowcon.com](mailto:technicalsupport@crowcon.com)

