

# Crowcon Technical Note

**Document Reference: GEN094**  
**Document applies to: T3 and Gasman**  
**Release Date: 24/10/2019**  
**Subject: HCN sensor iModule configurations in Gasman & T3**



We are making the following changes to the configuration of our HCN iModules with immediate effect:

- HCN iModule range was 0-30ppm - **now 0-25ppm**
- HCN iModule alarm level 1 is unchanged at 5ppm
- HCN iModule alarm level 2 was 25ppm - **now 10ppm**

These changes are more aligned to customer usage patterns and now match Cal cert outputs

---

For more information please contact +44 (0)1235 557711, [technicalsupport@crowcon.com](mailto:technicalsupport@crowcon.com)